

# UltraVac+

Vacuum/Drying Chamber



## Hearing Instrument Restoration System

Designed to give dispensers and technicians a functional tool for repairing, restoring and maintaining hearing aids.

This upgraded system is small and effectively removes cerumen and debris from hearing aid tubing and ports. It offers vacuum and positive air flow simultaneously to get even the most stubborn wax out of hearing aids.

### MedRx ULTRA VAC

- Digital Display of Chamber Pressure and Time Remaining
- Automatic Detection of Operating Mode
- Separate Pressure and Vacuum Wands Minimize Debris in System
- In-Line Vacuum Filter for Added Reliability
- 4 Minute Drying Chamber
- Auto Shut Off

# ULTRA VAC+

**MedRx, Inc**  
1200 Starkey Rd., Suite 105,  
Largo, FL 33771  
727-584-9600  
Toll Free: 888-392-1234  
Fax: 727-584-9602  
Email: sales@medrx-usa.com  
Web: www.medrx-usa.com

# Welcome to the New Generation.

**Powerful, small and  
effective hearing  
instrument restoration  
system**

## ULTRA VAC

### About MedRx

MedRx, Inc. is a U.S. based global manufacturer and developer of advanced computerized diagnostic and hearing instrument fitting technologies, specifically designed for the hearing care professional.

MedRx has created a remarkable New Generation of discreet, yet powerful, PC-based instrumentation for Audiometry, Real Ear Measurement, Live Speech Mapping, Hearing Instrument Testing & Evaluation and Video Otoscopy.

### MedRx UltraVac Accessories

- Durable Vacuum & Pressure Wands
- Assorted Wand Tips
- 4 Piece Toolkit
- Drying Chamber Cup
- Debris Filter

### Technical Specifications

**POWER REQUIREMENTS**  
100 - 240, 50/60 Hz

**POWER CONSUMPTION**  
<100 watts

**FUSE**  
1.25 amp SB 5 x 20 mm

**OPERATING TEMPERATURE**  
0° to 50° C (32° to 120° F)

**WEIGHT**  
Approx. 11 lbs • 5 kg

**DIMENSIONS**  
Approx 11" x 10.5" x 6.5" (L x W x H)  
Approx 28 cm x 26 cm x 16 cm (L x W x H)

**PARTICULATE FILTER**  
I-35S

**VACUUM TIPS**  
#14, #15, 2-#18 and 2-#20 gauge  
w/Safety Lock attachment

**DRYING CHAMBER CYCLE TIME**  
4 minutes with auto shut-off

**VACUUM WAND**  
Constant

